

Bid Document

Bid Details	
Bid End Date/Time	23-12-2019 15:00:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	70 (Days)
Ministry/State Name	Puducherry
Department Name	Higher And Technical Education Department Puducherry
Organisation Name	Tagore Arts College (tac) Pondicherry University
Office Name	Office Of The Principal
Total Quantity	20
Item Category	blood pressure recording units
Experience Criteria	1 Year (s)
MSE Exemption For Years Of Experience And Turnover	No
Startup Exemption For Years Of Experience And Turnover	No
Document required from seller	Experience Criteria,Past Performance,Certificate (Requested in ATC),OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	20 %
Bid to RA enabled	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 20% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Blood Pressure Recording Units (20 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
CONFIGURATION DETAILS	Type Of SPHYGMOM ANOMETERS or Blood Pressure recording unit:	Mercurial	Mercurial
	Conformity to Indian Standrad for SPHYG MOMANOMETERS	IS:3390 latest for Mercurial	IS:3390 latest for Mercurial
	Measuring device	Mechanical	Mechanical
	Type	Portable	Portable
	ISI MARKED	Yes	Yes
	CM/L No and validity date	-	*
	Measurement Method	Auscultatory method	Auscultatory method
	Display for Digital BP instrument:	NA for others	NA for others
	Range: Pressure measurement (mmHg)	0 to 300	*
	Range:Pulse measurement for digital with accuracy ± 5%. (per minute)	40 to 200	*
	Accuracy:Pressure	+/- 3	*

	measurement (mmHg)		
	Pressure Detection	Capacitive Pressure sensor	*
	Inflation	Manual	Manual
	Deflation and rapid Air Release	Manual	Manual
	Cuff Range	Medium (22-32)cm	Medium (22-32)cm
	Dimensions of the Equipment (mmXmm)	-	*
	Dimensions of the Display screen (mmXmm)	-	*
	Operating Temperature range	0 to 55 degree Celsius	*
	Storage Temperature range	(-10) to 55 degree Celsius	*
	Automatic power off if system is idle for 3 minutes.	Yes	*
	Weight (in gram)	800	*
Power	Power Source:	AA size chargeable battery set	AA size chargeable battery set
	Type of AA Size chargeable Battery (AA Size)	Alkaline	Alkaline
WARRANTY, CERTIFICATIONS, TEST REPORTS	Warranty Period (in Year)	1	1
	Availability of Type Test Report covering complete test parameters as per IS3390 latest for Mercurial and IS7652 latest for Aneroid type, from any ILAC/NABL accredited/Central Gov LABFor digital type including Environmental test from any ILAC/NABL accredited/Central Gov LAB	Yes	*
	Type test report No and date	-	*
	Name and address of the Lab where test Conducted.	-	*
	CE Marking Certificate	Yes	*

	for B.P. Apparatus Model.		
	Conformity to OILM-R-16-1 and OILM-R-16-2 for both Non-invasive mechanical sphygmomanometers and Non-invasive automated sphygmomanometers are for guidance and All the Gazette of india related to legal metrology issued time to time, are applicable to the specification as well as to the product.	Yes	*
	Others Certificate ie ROHS,WHO,GMP etc.	-	*
	For Digital Sphygmomanometer ,Availability of Type test reports consisting of verification of all the features & functional parameters & environmental tests sequences as under: a) Dry heat test (For 16 hrs at a temp of 55degree C in accordance with IS: 9000/part-3/section-3/1977)latest	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Nagaraj S	605008,Tagore Govt. Arts and Sceince College, Airport Road, Lawspet, Puducherry	20	15

Bid Specific Additional Terms and Conditions

1. OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be

ordered up to 25 percent at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods
3. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
4. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
5. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
6. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
7. ISO 9001: The bidder must have ISO 9001 certification.
8. Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---